

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



4613

Accredited to
ISO/IEC 17025:2017

Regroup (Reclaim) Limited

Issue No: 013 Issue date: 19 July 2019

Air Street
Bankside
Hull
East Yorkshire
HU5 1RR

Contact: Kira Allen
Tel: +44 (0)1482 879666
Fax: +44 (0)1482 879676
E-Mail: info@regroup.uk.com
Website: www.regroup.uk.com

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Petroleum and Petroleum Products Fuel Oil	<u>Chemical Tests</u> Carbon Residue Flash Point Strong Acid Number Total Sediment Kinematic viscosity at 40°C and 50°C Water Content <u>Sampling</u> Manual sampling	IP 13/94(14) BS 2000-13:94 ISO 6615:93 IP 523/15 BS 2000-523:15 BS EN ISO 3679:15 Rapid equilibrium closed cup method IP 139/98(17) BS 2000-139:98 ISO 6618:97 IP 375/11 BS 2000-375:11 ISO 10307-1:09 IP 71 Section 1/97 BS 2000-71:96 sect 1 BS EN ISO 3104:96 IP 74/2000(14) BS 2000-74:00 ISO 3733:1999 Documented in-house method SOP L54 based on IP 475/05 BS 2000-475:04 BS EN ISO 3170:04



4613

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Regroup (Reclaim) Limited

Issue No:013 Issue date: 19 July 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Burner Fuels Derived from Waste Mineral Oils	<u>Chemical Tests</u> Chlorine content Determination of Metals: Pb, Ni, Cr, Cu, Zn, As, Cd, Tl, Sb, Co, Mn and V Determination of Hg Sulphated Ash Sulphur content	 IP 503/04 BS 2000-503:04 ISO 15597:01 using in house method SOP L71 and WDXRF IP 593/11 using in house method SOP L70 and WDXRF IP608/15 using in house SOP L85 and WDXRF IP 550/08 ASTM D2622/16 using in house method SOP L73 and WDXRF
END		